



ALLIANCE™

(<https://www.globalseafood.org>).

---



 Responsibility

---

# Accuracy of custom water analyses varies

1 May 2013

By Phuong Thuy T. Le and Claude E. Boyd, Ph.D.

**pH, dissolved oxygen, total ammonia nitrogen and salinity are key parameters**



Lab equipment that utilizes spectrophotometry can simultaneously measure multiple elements at a reasonable cost, but results may not be consistently reliable.

Water quality data often are necessary for assessing the suitability of water supplies and water in culture units for fish, shrimp or other species. Analyses of pH, dissolved oxygen, total ammonia nitrogen and salinity can easily be conducted on site with hand-held instruments or water analysis kits, but data on concentrations of major ions and trace elements normally are obtained by sending samples to custom laboratories.

These laboratories usually analyze samples using atomic absorption spectrophotometry or, increasingly, by inductively coupled plasma atomic emission spectrophotometry (ICP-AES). The ICP-AES method typically provides data on concentrations of 16 to 20 elements measured simultaneously at a reasonable cost. Of course, alkalinity (bicarbonate plus carbonate) is measured by titration with a standard solution of acid.

## Ion concentration

Most laboratories have effective quality control programs to assure the reliability of results. Nevertheless, in the authors' experience, their reports on concentrations of major ions often are in error. Upon receiving the results for a water sample, it is possible to assess the probable reliability of the data. The usual procedure is to make an anion-cation balance evaluation.

Because of the principle of electrical neutrality, positive charges of cations balance negative charges of anions in a sample. The major ions typically account for more than 95 percent of the total ions in a sample. Thus, if the concentrations of major ions are converted from a weight basis (milligrams per liter) to a charge basis (milliequivalents per liter), the sum of the milliequivalents of cations (calcium, magnesium, sodium and potassium) very nearly equals the sum of the milliequivalents of anions (bicarbonate/carbonate, chloride and sulfate) in an accurate analysis.



(<https://www.deviseafoods.com>).

## Working example

Suppose a sample is reported to contain 40 mg/L calcium, 12 mg/L magnesium, 10 mg/L sodium, 4 mg/L potassium, 120 mg/L bicarbonate, 33 mg/L chloride and 24 mg/L sulfate. The charge balance can be made as follows:

### Anions

120 mg/L bicarbonate ÷ 61 mg bicarbonate/meq = 2.00 meq/L

33 mg/L chloride ÷ 35.45 mg chloride/meq = 0.93 meq/L

24 mg/L sulfate ÷ 48 mg sulfate/meq = 0.50 meq/L

Total negative charge = 3.43 meq/L

### Cations

40 mg/L calcium ÷ 20.04 mg calcium/meq = 2.00 meq/L

12 mg/L magnesium ÷ 12.15 mg magnesium/meq = 0.99 meq/L

10 mg/L sodium ÷ 23 mg sodium/meq = 0.43 meq/L

4.0 mg/L potassium ÷ 39.1 mg potassium/meq = 0.10 meq/L

Total positive charge = 3.52 meq/L

The charge balance (meq anions ÷ meq cations) in the example is almost equal – 0.974 as compared to 1.0 for perfect balance – suggesting an accurate analysis. The charge balance, however, is not always proof of a reliable analysis.

For example, a sample of seawater the authors recently sent to a laboratory for analysis by ICP-AES was reported to contain 164 mg/L bicarbonate; 26,552 mg/L chloride; 3,763 mg/L sulfate; 575 mg/L calcium; 1,790 mg/L magnesium; 14,858 mg/L sodium and 733 mg/L potassium. These concentrations were all greater than would be expected for normal seawater. The sample was reported to have a conductivity of 50,300  $\mu$ mhos/cm, a typical conductivity for normal seawater.

A check with a hand-held salinometer when the sample was taken gave a salinity of 36 ppt, a reasonable estimate for normal seawater. However, the reported concentrations of most ions were higher than expected for normal seawater, and the sum of the ions was 48,435 mg/L – normal seawater should contain about 35,000 mg/L of total ions. Thus, in spite of the ion balance being very good (0.987), the analysis was obviously in error.

To resolve the issue with the seawater sample, the major ions were measured by traditional methods with the following results: bicarbonate, 126 mg/L; chloride, 20,660 mg/L; sulfate, 2,677 mg/L; calcium 340 mg/L; magnesium, 1,276 mg/L; sodium, 11,260 mg/L; potassium 412 mg/L. The sum of these ions was 36,750 mg/L, and the specific conductance was 49,800  $\mu$ mhos/cm – reasonable values for normal seawater.

For seawater and brackish water samples, conductivity multiplied by 0.7 is a relatively good estimate of total dissolved solids. In the seawater example above, the estimate of total ion concentration based on the reported conductivity of 50,300  $\mu\text{mhos/cm}$  was 35,210 mg/L. This simple calculation could have alerted the analyst to the fact that the analysis was in error, because the concentrations of the major ions measured by ICP-AES totaled 48,435 mg/L.

The factor 0.7 is not as reliable for estimating the total dissolved-solids concentration from conductivity in many freshwater samples as it is for seawater and brackish water samples. The factor has been reported to range from 0.55 to 0.80 for freshwater. Nevertheless, the factor of 0.7 usually gives an indication of whether a particular analysis of freshwater might be in error despite having a satisfactory charge balance.

## Synthetic water test

To assess the reliability of data on major ion concentrations measured by custom laboratories, the authors made a synthetic seawater sample, analyzed it by traditional methods and sent it to five custom laboratories (Table 1). The synthetic seawater had a measured conductivity of 49,800  $\mu\text{mhos}$ , a reasonable expectation for normal seawater. The traditional methods of analysis provided ionic concentrations similar to expected concentrations, the charge balance was 1.03, and the total concentration of measured ions was similar to the estimated total ion concentration.

## Boyd, Results of analyses of major ions in synthetic seawater, Table 1

Ion	Weighed Amount	Traditional Analysis	Laboratory 1	Laboratory 2	Laboratory 3	Laboratory 4	L
Bicarbonate (mg/L)	145	136	117	129	145	179	
Chloride (mg/L)	19,488	20,660	20,000	22,179	19,500	5,229	
Sulfate (mg/L)	2,719	2,677	2,700	3,077	2,698	2,464	
Calcium (mg/L)	409	340	694	404	258	411	
Magnesium (mg/L)	1,304	1,276	2,238	1,352	1,266	1,199	
Sodium (mg/L)	10,853	11,260	27,487	10,172	9,343	10,050	
Potassium (mg/L)	383	412	1,045	533	508	405	
Total (mg/L)	35,301	36,761	54,261	37,846	33,718	19,937	
Conductivity ( $\mu\text{mhos/cm}$ )	–	49,800	41,540	47,900	49,800	55,279	
Total ions, estimated (mg/L)	–	34,860	29,078	33,530	34,860	38,695	

Anions ÷ cations	0.999	1.030	0.431	1.178	1.135	0.356
Total <sub>m</sub> ÷ total <sub>e</sub>	–	1.055	1.866	1.129	0.967	0.515

Total<sub>m</sub> = Measured total (sum of ions). Total<sub>e</sub> = Specific conductance x 0.7

Table 1. Results of analyses of major ions in synthetic seawater measured by traditional methods and custom laboratories using ICP-AES.

The custom laboratories reported conductivities ranging from 41,540 to 55,279 µmhos. The range in concentrations of individual cations was huge – 361-1,045 mg/L for potassium, 258-694 mg/L for calcium and 5,229-22,179 mg/L for chloride. The charge balances for the results from the five laboratories ranged from 0.349 to 1.178, while the ratio of total measured ion concentration to estimated total ion concentration varied from 0.515 to 1.866. Only two of the laboratories provided results that reflected the actual composition of the sample.

## Reliability of results

Although these findings do not bode well for the accuracy of ICP-AES analyses conducted by laboratories, most custom laboratories mainly run samples of a particular type – soil samples in the case of the laboratories used to test the synthetic water. In spite of the fact that these laboratories also do routine water analyses, they may not always adjust the ICP-AES methodology accordingly.

In fact, laboratory 1 that provided fairly reasonable results in Table 1 was the same laboratory that had earlier provided erroneously high results for the seawater sample discussed above. Before sending the synthetic seawater sample, the problem of the earlier seawater analysis was discussed with the lab director. Despite his icy response during the conversation, he apparently took heed, for the laboratory performed much better on the synthetic seawater sample.

*(Editor's Note: This article was originally published in the May/June 2013 print edition of the Global Aquaculture Advocate.)*

## Authors



**PHUONG THUY T. LE**

Department of Fisheries and Allied Aquacultures  
Auburn University



**CLAUDE E. BOYD, PH.D.**

Department of Fisheries and Allied Aquacultures  
Auburn University  
Auburn, Alabama 36849 USA

[boydce1@auburn.edu](mailto:boydce1@auburn.edu) (<mailto:boydce1@auburn.edu>).

Copyright © 2023 Global Seafood Alliance

All rights reserved.